



## Recommended Due Date for Calibration

Model: SDM3055

Serial NO.: SDM35FAC4R 0024

*SIGLENT has determined that the factory calibration of this instrument is not significantly affected by storage of up to 180 days before first-time use. Cal Interval should start at the time the unit is placed in service or 180 days past the "Date of Calibration" on the certificate received with the unit.*

Determine the date when the next cal is due (refer to the examples shown below)

**Example 1:** *First-time use of the instrument is **less than** 180 days after DATE-TESTED.*

Date tested at Factory: June 17, 2019

Example Calibration Interval: 1 Year

First-Time Use of Instrument: September 2, 2019

First-Time Use + Calibration Interval = Date for Next Calibration

---

September 2, 2019 + 1 Year = September 2, 2020

**Example 2:** *First-time use of the instrument is **more than** 180 days after DATE-TESTED.*

Date tested at Factory: March 22, 2019

Example Calibration Interval: 1 Year

First-Time Use of Instrument: October 10, 2019

Date Tested at Factory + 180 Days + Calibration Interval = Date for Next Calibration

---

March 22, 2019 + 180 Days + Calibration Interval = September 19, 2020

Checked by:

*Yan Wang*