

Demonstration of the SIGIQPro Bluetooth signal generation and SSA5000A Bluetooth analysis function

February 02, 2024

1 Introduction

As a wireless communication technology, Bluetooth technology is widely used in various devices and application scenarios. The article aims at RF verification of Bluetooth transmitter. Through fast one-click RF measurement, SSA5000A spectrum analyzer is transformed into a standard-based Bluetooth RF transmission tester to help you design, evaluate and manufacture Bluetooth devices. The measurement application conforms to the standard of Bluetooth core specification, and can safely verify your Bluetooth design, covering Bluetooth BR, EDR and LE. The article will demonstrate how to use the Bluetooth analysis function of SSA5000A to carry out Bluetooth analysis and measurement quickly and effectively.

1.1 Bluetooth Analysis Function

Bluetooth SIG specifies the RF test items of Bluetooth® Classic and Bluetooth® Low Energy Bluetooth measurement specifications. Tables 2 and 3 list the corresponding test items supported by SSA5000A and suitable for transmitter testing. The SIGLENT Bluetooth measurement application refers to the following Bluetooth RF test specifications:

Table 1 Supported standard versions

Туре	Bluetooth device			
	Bluetooth radio frequency system specification 1.2/2.0/2.0+EDR/2.1/2.1+EDR revision 2.1.E.0			
Standard	— basic rate			
version	 enhanced data rate Bluetooth Low Energy RE PHY test specification (LE RE-PHY TS/4 0/4 2) 			
	Bluetooth RF PHY test specification (Bluetooth 5, 5.1 and 5.3)			
Power	Class 1, class 2 and class 3			
	Bluetooth basic rate and EDR system:			
Frequency	2.400 to 2.4835 GHz (f = 2402 +k MHz, k = 0,, 78) Bluetooth low energy system:			
	2.400 to 2.4835 GHz (f = 2402 +kx2 MHz, k = 0,, 39)			
	Table 2 Supported Classic test items			
Transmitter tes	st TP/TRM/CA/BV-xx-C			
Basic Rate (BR				
Output power	01			
Modulation cha	aracteristic 07			
Initial carrier fr	equency tolerance 08			
Carrier frequer	ncy drift 09			

Enhanced Data Rate (EDR)



EDR relative transmit power 10 EDR carrier frequency stability and 11 modulation accuracy EDR differential phase encoding 12 Table 3 Supported LE test items Low Energy (LE) LE 1M LE 2M LE 1M LE 2M Coded, S=8 Transmitter test 1 Mb/s 2 Mb/s 1 Mb/s, SMI 2 Mb/s, SMI 1 Mb/s, SMI Output power 01 Modulation characteristics 05 10 09 11 13 Carrier frequency offset and drift 06 12 14

1.2 Configuration information

You need to install the appropriate firmware version and options on the required instruments.

Table 4 Configuration information

Туре	Model	Configuration information
Signal Generation Software	SigIQPro	Optional: SigIQPro-BT
Vector Signal Generator	SSG5000X-V	Firmware: V2.1.2.4.1and above Optional: SigIQPro-BT
Arbitrary Waveform Generator	SDG7000A	Firmware: V1.1.1.32 and above Optional: SigIQPro-BT
Spectrum Analyzer	SSA5000A	Firmware: V1.1.2.2.0 and above Optional: SSA5000A-BT

1.3 Measurement Parameters

SSA5000A Bluetooth analysis provides one-button measurement.

Bluetooth Analysis	Basic Rate (BR)	Enhanced Data Rate (EDR)	Low Energy (LE)			
Output power						
u Peak power	•		•			
u Average power	•		•			
Modulation characteristics						
u ∆F1 Avg/Max	•		•			
u ΔF2 Avg/Max	•		•			
u Δ F1/ Δ F2 Ratio	•		•			



Initial carrier frequency tolerance (ICFT)		
Frequency offset		•
Carrier frequency drift		
u Frequency drift		•
u Max drift rate		•
u Init freq drift		•
Relative transmit power		
u GFSK average power	•	
u DPSK average power	•	
u Rel average power	•	
Frequency stability and modulation	n accuracy	
u Freq offset	•	
u Freq offset	•	
u +	•	
u RMS DEVM	•	
u Peak DEVM	•	
Differential phase decoding		
u BER	•	
u Bit errors	•	
u 99% DEVM	•	
Guard Interval	•	

1.4 Connection Settings

Connect a PC with SigIQPro software to SSG5000X-V / SDG7000A through GPIB / LAN / USB. Complete the connection according to the setup instructions of SigIQPro, and then perform the following steps to interconnect SSG5000X-V / SDG7000A and SSA5000A:

1) Connect SSG5000X-V/ SDG7000A RF output port to SSA5000A RF input port.

It is suggested to connect the 10 MHz IN output of SSG5000X-V / SDG7000A to the Ext RefI port (back panel) of SSA5000A to improve the frequency accuracy

SIGLENT[®]



Figure 1-1 Connection settings

1.5 Test Result

SSA5000A calculates the measurement results according to the Bluetooth® RF test specification, carries out Bluetooth BR/ EDR/LE one-button measurement and displays a single view with four traces. The one-button measurement function can conveniently and quickly complete the test purposes listed in Table 2 and 3 at one time, and provide an overview of multiple emission tests. Users can switch from the four-trace overview to a single power, modulation and spectrum measurement display to view the results in more detail. All signal parameters can be modified independently, such as RF channel, packet type, mode type, etc. SSA5000A is helpful for users to troubleshoot flexibly in the laboratory and optimize RF design easily and quickly.

SIGLENT[®]



Figure 1-2 Test result



North American Headquarters

SIGLENT Technologies NA 6557 Cochran Rd Solon, Ohio 44139 Tel: 440-398-5800 Toll Free:877-515-5551 Fax: 440-399-1211 info@siglent.com www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES GERMANY GmbH Staetzlinger Str. 70 86165 Augsburg, Germany Tel: +49(0)-821-666 0 111 0 Fax: +49(0)-821-666 0 111 22 info-eu@siglent.com www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD. Blog No.4 & No.5, Antongda Industrial Zone, 3rd Liuxian Road, Bao'an District, Shenzhen, 518101, China. Tel:+ 86 755 3661 5186 Fax:+ 86 755 3359 1582 sales@siglent.com www.siglent.com/ens