

Revision History

Version	Revision
V1.0.3.1.21 2023/09/6	<ol style="list-style-type: none">Fixed the defects related to the flatness list functionFixed the defects related to SCPI commandFixed the defects related to the VNC function.
V1.0.3.1.19R5 2020/04/28	<ol style="list-style-type: none">Fix a bug for power meter functionFix a bug for power meter filling flatness
V1.0.3.1.19R2 2019/04/16	<ol style="list-style-type: none">Add the function of flatness, new version provide two types to fill the flatness table. One is automatically with the power sensor or according to set the frequency step or the points to fill the flatness tableAdd function of adjust reference accuracySolution some bugs of pulse train functionOptimize the AM performanceSolution some bugs of the FM functionSolution some bugs of the VNC control
V1.0.3.1.18 2018/11/1	<ol style="list-style-type: none">Add option of upgrade BW from 2.1GHz to 3.2GHzOptimize UI operationOptimize the state fileOptimize the IP state symbolOptimize the presetOptimize the LF sweep trigger

Update Instructions

1 **Note: During the updating, do not cut off the power or pull out the U-disk.**

1. Please firstly copy the update file (.ADS) to a U-disk.
2. Insert the U-disk.
3. Press **UTILITY**, we can enter **Update** in **UTILITY** menu.




Figure 1 UTILITY function menu

4. Press **Update** and then enter the file system (or you can press **Store/Recall** button to enter the file system).

U-disk0:/V30R005S03D01P18.ADS				Size:559.8MB/7.3GB
Name:	Size:	Type:	Date:	
U-disk0		Folder	1970/01/01	
V30R005S03D01P18.A...	8.6 MB	ADS	2018/10/29	
V30R005S03D01P17r1....	25.8 MB	ADS	2018/10/02	
V30R005S03D01P17.ADS	5.5 MB	ADS	2018/10/02	
V30R005S03D01P16r2....	5.5 MB	ADS	2018/10/02	
V30R005S03D01P16r1....	29.3 MB	ADS	2018/07/31	
V30R005S03D01P16.A...	29.3 MB	ADS	2018/07/24	
V30R005S03D01P15r2....	30.6 MB	ADS	2018/07/10	
V30R005S03D01P15r1....	30.6 MB	ADS	2018/06/27	
V30R005S03D01P15.A...	27.5 MB	ADS	2018/06/12	

Recall **New Dir** **Delete** **Rename** **Copy** **Return**

Figure 2 File system

5. Select the update firmware which should be with an ADS postfix. Press **Recall** to start the update. A progress bar is displayed during the update. After the update is completed, the generator will restart automatically.
6. After the signal generator restarted, we can check the version information in System Info.