SDS5000X Security Erase Process

February 12, 2019

In some instances, you may wish to clear all of the available memory on an instrument to prevent anyone from seeing the data or information.

Here are the steps to securely erase the memory modules within the SIGLENT SDS5000X Oscilloscope Series:

Press Utility > Save/Recall > Recall > Type = Security Erase > Press Recall


Analysis
SIGLENT
Auto
f < 2.0Hz
RECALL
Setup
Reference
Factory Default
Security Erase
Mode
Save
Recall
Type
Security Erase
Recall
Timebase
0.00s
2.00us/div
100kpts
5.00GSa/s
Trigger
C1 DC
Auto
0.00V
Edge
Rising
14:20:09
2019/2/12

North American Headquarters

SIGLENT Technologies NA
6557 Cochran Rd Solon, Ohio 44139
Tel: 440-398-5800
Toll Free:877-515-5551
Fax: 440-399-1211
info@siglent.com
www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES GERMANY GmbH
Staetzlinger Str. 70
86165 Augsburg, Germany
Tel: +49(0)-821-666 0 111 0
Fax: +49(0)-821-666 0 111 22
info-eu@siglent.com
www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD.
Blog No.4 & No.5, Antongda Industrial Zone,
3rd Liuxian Road, Bao'an District,
Shenzhen, 518101, China.
Tel:+ 86 755 3661 5186
Fax:+ 86 755 3359 1582
sales@siglent.com
www.siglent.com/ens