

What is the largest datalogging file I can store with the XE series oscilloscopes?

December 13, 2021

The SIGLENT SDS1104X-E and SDS1204X-E oscilloscopes have a data-logging feature that was added in [firmware revision 6.1.37R2 and OS Rev 2](#)

As of now, the maximum file storage size for a data logging event is 2 GB and the measurement logger is 14 Mpts. The USB stick must be as big, or bigger than 2 GB to reach this, of course. The logging process will terminate when this limit is reached.

Learn more by clicking on this [Datalogging Application Note](#)



North American Headquarters

SIGLENT Technologies NA
6557 Cochran Rd Solon, Ohio 44139
Tel: 440-398-5800
Toll Free: 877-515-5551
Fax: 440-399-1211
info@siglent.com
www.siglentamerica.com/

European Sales Offices

SIGLENT TECHNOLOGIES GERMANY GmbH
Staetzlinger Str. 70
86165 Augsburg, Germany
Tel: +49(0)-821-666 0 111 0
Fax: +49(0)-821-666 0 111 22
info-eu@siglent.com
www.siglenteu.com

Asian Headquarters

SIGLENT TECHNOLOGIES CO., LTD.
Blog No.4 & No.5, Antongda Industrial Zone,
3rd Liuxian Road, Bao'an District,
Shenzhen, 518101, China.
Tel: + 86 755 3661 5186
Fax: + 86 755 3359 1582
sales@siglent.com
www.siglent.com/ens